



HEAT SOURCE ACCORDING TO DIN EN 12722

- Ø cone surface area 100 mm, height 55 mm.
 - centric drilling Ø 6,5 mm to the measurement of the temperature (margin of error $\pm 1^{\circ}$ C).
-
- **Testing Method:** A standard block from aluminium according to ISO 209:2007 with a fixed inspection temperature is placed on a surface and removed after a certain time again.

SKU: 04800-A

Categories: [Test Equipment](#), [Furniture Testings](#), [Surface loading device](#)

