



## HEAT SOURCE ACCORDING TO DIN EN 12722

- $\varnothing$  cone surface area 100 mm, height 55 mm.
- centric drilling  $\varnothing$  6,5 mm to the measurement of the temperature (margin of error  $\pm 1^\circ$  C).
  
- **Testing Method:** A standard block from aluminium according to ISO 209:2007 with a fixed inspection temperature is placed on a surface and removed after a certain time again.

---

**SKU:** 04800-A

**Categories:** [Test Equipment](#), [Furniture Testings](#), [Surface loading device](#)

