

## Electrical Safety

### Safety Finger Probe according to VDE 0470

pivot finger probe  $\varnothing$  12 mm and 80 mm length  
sufficient distance from the dangerous parts.

**Art. No. # 11300-A**



fixed finger probe  
with force scale 0 - 100 N  
according VDE.

**Art. No. # 11300-B**



fixed finger probe 4 / 3 x 15 mm.

**Art. No. # 11300-C**

